



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: July 20, 2010 INITIALS: Sec OWNER: WASHINGTON STATE PATROL
DISTRICT: 7 ADDRESS: I-5 MP235
LOCATION: BowHill 33 E1 COUNTY: Skagit
MAKE: Toledo TYPE Trk Scale ACCURACY CLASS: III L
MODEL: 8530 ID. NO: 4409217-4SX CAPACITY: 100000 lbs. SECTIONS: 2
MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 12x20

East 1

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
SEC 1	20000	20000	0				
CENTER	20000	20000	0				
SEC 2	20000	20020	+20				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-2 and 200446-0-L2902-8

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer

Date: Aug 4, 2010

Print name: David Cromer

State of Washington
County of Shunston
Signed and sworn to (or affirmed before
me on Aug 4, 2010 by

Brenda A Biscay
(Signature of notary public)
Brenda A. Biscay
Printed Name of Notary
My appointment expires 3.12.12

